# QUICK GUIDE

VIDEO TUTORIALS




206.1179.0

## SCL / SVCL

# CLOSED AMPEROMETRIC CELL

## BODY




**IT**

**A** Elettrodo
**B** Micro Foro
**C** Cappuccio Trasparente
**D** Tappo a Membrana

**ES**

**A** Electrodo
**B** Microagujero
**C** Tapa transparente
**D** Tapa de membrana

**EN**

**A** Electrode
**B** Micro Hole
**C** Transparent Cap
**D** Membrane cap

**FR**

**A** Électrode
**B** Micro-trou
**C** Capuchon transparent
**D** Capuchon à membrane

**DE**

**A** Elektrode
**B** Mikroloch
**C** Transparente Kappe
**D** Membrankappe

**PL**

**A** Elektroda
**B** Mikrootwór
**C** Przezroczysta nasadka
**D** Nakładka membranowa

# CLOSED AMPEROMETRIC CELL

## BUFFER SOLUTION SETUP

**UTILIZZARE I DISPOSITIVI DI PROTEZIONE INDIVIDUALE**
**USE PERSONAL PROTECTIVE EQUIPMENT**
**VERWENDEN SIE PERSÖNLICHE SCHUTZAUSRÜSTUNG**
**UTILICE EQUIPO DE PROTECCIÓN PERSONAL**
**UTILISER UN ÉQUIPEMENT DE PROTECTION INDIVIDUELLE**
**STOSOWAĆ ŚRODKI OCHRONY INDYWIDUALNEJ**

### SCL SVCL / 3S / 8




### SCL SVCL 3N / 11




### SCL SVCL 9




### SCL SVCL 2




**A**
Cappuccio Sonda
Probe cap
Sondenkappe
Tapa de sonda
Capuchon de sonde
Nasadka sondy

**B**
Livello rimpimento elettrolita
Electrolyte filling level
Füllstand Elektrolyt
Nivel de llenado de electrolito
Niveau de remplissage d'électrolyte
Poziom napełnienia elektrolitem

**C**
Pipetta
Pipette
Pipette
Pipeta
Pipette
Pipeta

**D**
Foro di ventilazione
Ventilation Hole
Belüftungsloch
Orificio de ventilación
Trou d'aération
Otwór wentylacyjny

# CLOSED AMPEROMETRIC CELL

## ON “PEF” INSTALLATION

**UTILIZZARE I DISPOSITIVI DI PROTEZIONE INDIVIDUALE**
**USE PERSONAL PROTECTIVE EQUIPMENT**
**VERWENDEN SIE PERSÖNLICHE SCHUTZAUSRÜSTUNG**
**UTILICE EQUIPO DE PROTECCIÓN PERSONAL**
**UTILISER UN ÉQUIPEMENT DE PROTECTION INDIVIDUELLE**
**STOSOWAĆ ŚRODKI OCHRONY INDYWIDUALNEJ**

**A**
Anello Posizionamento Sonda
Probe Positioning Ring
Sondenpositionierungsring
Anillo de posicionamiento de la sonda
Bague de positionnement de la sonde
Pierścień pozycjonujący sondę

**B**
Anello Fissaggio Sonda
Probe Fixing Ring
Sondenbefestigungsring
Anillo de fijación de sonda
Anneau de fixation de sonde
Pierścień mocujący sondę

**C**
Oring tenuta sonda
Probe sealing O-ring
Sondendichtungs-O-Ring
Junta tórica de sellado de sonda
Joint torique d'étanchéité de la sonde
O-ring uszczelniający sondę

**D**
Anello porta Oring
O-ring holder ring
O-Ring-Haltering
Anillo de soporte de junta tórica
Anneau de support de joint torique
Pierścień uchwytu o-ringu

**E**
Oring tenuta PEF
PEF sealing O-ring
PEF-Dichtungs-O-Ring
Junta tórica de sellado pef
Joint torique d'étanchéité pef
Oring uszczelniający pef

**VERIFICARE LA CORRETTA ISTALLAZIONE DELLA SONDA (1, 2)**
**CHECK FOR PROPER CELL INSTALLATION (1, 2)**
**PRÜFEN SIE AUF RICHTIGEN ZELLEINSTALLATION (1, 2)**
**COMPRUEBE QUE LA CELDA ESTÁ INSTALADA CORRECTAMENTE (1, 2)**
**VÉRIFIER LA BONNE INSTALLATION DES SONDE (1, 2)**
**SPRAWDŹ PRAWIDŁOWĄ INSTALACJĘ SONDY (1, 2)**









# CLOSED AMPEROMETRIC CELL

## SCL CABLE CONNECTION

**OLD CABLE** | **CONTACTS**
SCL PROBE

- → (-485)
- → (+485)
- → (GND)
- → (+5 VDC)



GREY CABLE (OLD)

**NEW CABLE** | **CONTACTS**
SCL PROBE

- → (-485)
- → (+485)
- → (GND)
- → (+5 VDC)
- *Not Connected*

GREEN CABLE (NEW)

# CLOSED AMPEROMETRIC CELL

## SVCL CABLE CONNECTION

**OLD CABLE** | **CONTACTS**
SVCL PROBE

- → (V+)
- → (V-)
- → (IN)
- → (GND)

**NEW CABLE** | **CONTACTS**
SVCL PROBE

- → (V+)
- → (V-)
- → (IN)
- → (GND)

DOWNLOAD MANUAL